

DCS™

Plantar Fasciitis Sleeve



DCS-PKG-21_Rev0321



EC REP
DARCO® (Europe) GmbH
Gewerbegebiet 18
82399 Raisting, Germany
www.darco.de
Made in China

Manufactured for:
DARCO® International
810 Memorial Blvd.
Huntington, WV 25701 | USA
Toll Free: 800-999-8866
Phone: 304-522-4883
www.darcointernational.com

Instructions for use located on inside.

Available in 3 colors:



Size:	Arch Circ.	Women:	Men:	Blue:	Black:	Tan:
S	5 - 8 in	4 - 6.5	3 - 5.5	DCS-PF1	DCS-PF1-B	DCS-PF1-T
M	7 - 10 in	7 - 10.5	6 - 9.5	DCS-PF2	DCS-PF2-B	DCS-PF2-T
L	9 - 13 in	11+	10 - 13	DCS-PF3	DCS-PF3-B	DCS-PF3-T
XL	11 - 15 in	-	13.5+	DCS-PF4	DCS-PF4-B	DCS-PF4-T

DCS™ Sizing (limited U.S.A. availability)

- › The DARCO® DCS™ Plantar Fasciitis Sleeve is woven to be anatomically correct. Wave compression technology on the plantar surface of the foot provides a sustained stretch on the plantar fascia. The DCS™ gives relief from the discomfort associated with plantar fasciitis by promoting circulation, reducing edema and foot
- › Wave compression technology provides a sustained stretch to the plantar fascia.
- › Comfort Zone - Special knitting pattern at the Anterior Ankle Fold to help prevent pinching and binding.
- › 24 hour support - wear while sleeping or under socks when active.
- › Constructed of a medical grade fabric that is latex free.
- › Anatomically designed for better comfort.

Indications

> Plantar fasciitis, heel spurs, arch pain, arch weakness, edema

Contraindications

> Poor circulation, diabetic foot/neuropathy

Cleaning Instructions/Maintenance

> When the DCS™ is slightly dirty, clean it with a damp cloth or a soft brush.
> Allow the DCS™ to air dry only.

Instructions for use

> The top opening is indicated with the DARCO® logo.
> Pull the sleeve over the foot until the heel is properly positioned in the heel pocket.

Caution

Please always follow these instructions for use. In case of pain or swelling, discontinue use and consult your medical provider immediately. This device is designed to assist recovery and should be used exclusively under the instruction of and in consultation with a healthcare professional. DARCO® International Inc. does not assume any responsibility for misuse of the product. Dressings or socks can be worn with the DCS™. Walk more slowly than normal. This device is for single-patient use only.

To the patient/user: This device has been designed and manufactured by DARCO® International Inc. to be state of the art for treatment of a specific medical condition. Should any serious incident occur in relation to this device it should be reported to the manufacturer at the contact information listed on this document and the proper authority of the country or state where you are located. When treatment is complete, dispose of this device according to local laws and ordinances.

Indicaciones

> Fascitis plantar, espolones calcáneos, dolor en el arco, debilidad del arco, edema

Contraindicaciones

> Mala circulación, pie diabético/neuropatía

Instrucciones de limpieza/mantenimiento

> Cuando el DCS™ esté un poco sucio, límpielo con un paño húmedo o un cepillo blando.
> Seque el DCS™ únicamente al aire libre.

Instrucciones de uso

> La abertura superior está indicada con el logotipo de DARCO®.
> Pase la manga por encima del pie hasta que el talón esté correctamente colocado en el bolsillo del talón.

Precaución

Siga siempre estas instrucciones de uso. En caso de dolor o hinchazón, deje de usarlo y consulte a su proveedor médico de inmediato. Este dispositivo está diseñado para ayudar en la recuperación y debe usarse exclusivamente bajo las instrucciones de un profesional de la salud, luego de una consulta. DARCO® International Inc. no asume ninguna responsabilidad por el mal uso del producto. Se pueden usar vendajes o calcetines con el DCS™. Camine más despacio de lo normal. Este dispositivo es para uso exclusivo de un solo paciente.

Para el paciente/usuario: este dispositivo ha sido diseñado y fabricado por DARCO® International Inc. con el fin de otorgar la calidad más avanzada en el tratamiento de una afección médica específica. Si ocurriera algún incidente grave en relación con este dispositivo, se debe informar al fabricante en la información de contacto que figura en este documento y a la autoridad correspondiente del país o estado donde se encuentra. Una vez finalizado el tratamiento, deseché este dispositivo de acuerdo con las leyes y ordenanzas locales.

Indications

> Fasciite plantaire, épines calcanéennes, douleur à la voûte plantaire, faiblesse de la voûte plantaire, oedème

Contre-indications

> Mauvaise circulation, pied ou neuropathie diabétique

Instructions de nettoyage et entretien

> Si la DCS™ est légèrement sale, nettoyez-la avec un chiffon humide ou une brosse douce.
> Laissez la DCS™ sécher à l'air uniquement.

Mode d'emploi

> L'ouverture supérieure est indiquée par le logo DARCO®.
> Tirez la gaine sur le pied jusqu'à ce que le talon soit correctement positionné dans la poche du talon.

Mise en garde

Veuillez toujours suivre les présentes instructions d'utilisation. En cas de douleur ou d'enflure, cessez l'utilisation et consultez immédiatement votre médecin. Cet appareil est conçu pour aider à la récupération et doit être utilisé exclusivement sous les instructions d'un professionnel de la santé et en consultation avec celui-ci. DARCO® International Inc. n'assume aucune responsabilité en cas de mauvaise utilisation du produit. Des pansements ou des chaussettes peuvent être portés avec la DCS™. Marchez plus lentement que d'habitude. Cet appareil est destiné à un usage par un seul patient.

Pour le patient ou l'utilisateur : Cet appareil a la fine pointe de la technologie a été conçu et fabriqué par DARCO® International, Inc. pour le traitement d'une condition médicale spécifique. Si un incident grave survient relativement à cet appareil, il doit être signalé au fabricant aux coordonnées indiquées dans ce document et aux autorités compétentes du pays ou de l'État où vous vous trouvez. Une fois le traitement terminé, jetez cet appareil conformément aux lois et aux ordonnances locales.